

Weightron Bilanciai Ltd,
 Titan Works
 Bridge Way off Broombank Road,
 Chesterfield Trading Estate,
 Chesterfield S41 9QJ, UK
 Tel: 01246 260062
 Fax: 01246 260844
 www.weightron.com

Weightron Bilanciai Ltd,
 Unit 4b, Block 4
 Westpark, Chelston
 Wellington TA21 9AD, UK
 Tel: 01823 662355
 Fax: 01823 6621017
 www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the Calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Ardagh Glass Knottingley	
C/O Berrymans Headlands Lane Knottingley West Yorkshire WF11 0HP	
Location	Site

Machine Details

Cust ref	Weighbridge	
Make	Toledo	
Model	Cougar	
Serial No		
Capacity	50000	kg
e/d	20	kg
Working cap	44000	

Test Weights

Set No.	Cert No.	Date Cal
1-22	T#R12956	11/08/2026

Cert Interval	12 Months
Next Cert	Aug 2026

Tolerance +/-	kg
0 - 10000	20
10000 - 40000	40
40000 - 50000	60

Comments

--



AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
400	380	-20	380	-20
10000	10000	0	10000	0
18000	18000	0	18000	0
20920	20920	0	20920	0
38920	38920	0	38920	0
44220	44200	-20	44200	-20

DEFINITIVE LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev

Repeatability

Load kg	Ind	Ind	Ind
44220	44200	44200	44200

Repeatability

Load kg	Ind	Ind	Ind

Eccentric Load 8300

Reading	kg	Dev	Reading	kg	Dev
1	8280	-20	2	8280	-20
3	8300	0	4	8300	0
5	8300	0	6	8300	0
7			8		
9			10		
11			12		

Eccentric Load

Reading	kg	Dev	Reading	kg	Dev
1			2		
3			4		
5			6		
7			8		
9			10		
11			12		

Signature:

Name: Dave Waller

Date: 18/08/2025 10:19