

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

URM LTD KIDGATE CRESCENT SOUTH KIRBY WF9 3NR LOCATION SITE
--

Machine Details

Cust ref	WEIGHBRIDGE
Make	PT
Model	DD2050
Ser No	238905
Capacity	50/0001b
e/d	1201g
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
W31-22	TR12956	21.7.25

Cert Interval	YEARLY
Next Cert	AUGUST 2026

Tolerance +/-	Kg
0-10000	20
10000-40000	40
40000-50000	60

Comments *Make up weight for 5300g*
lorry 21160g

2	3
1	4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
9000	9000	0	400	0
18000	17980	20	9000	0
21160	21160	0	21160	0
39160	39140	20	39140	20
44460	44420	40	44420	40

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44460	44420	44440	44420

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 8300g

Reading	Dev	Reading	Dev
1 8300	0	2 8300	0
3 8300	0	4 8300	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature

Name *D. Waller*

Date *19.8.2025*